



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

DR. GOLIK – SCIENTIFIC SOLUTIONS  
34 Kibutz Galuyot St.  
Tel-Aviv, 6655018, ISRAEL  
Baruch Pilo Phone: 972 3 5187555

CALIBRATION

Valid To: November 30, 2022

Certificate Number: 5874.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,9</sup>:

I. Chemical

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments <sup>7</sup>
pH Meters <sup>3</sup> – pH	4 pH, 7 pH, 9 pH	0.021 pH	Standard pH solutions
Electrical Simulation of pH Indicator	Up to 14 pH	0.0022 pH	Simulator Metrohm

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments <sup>7</sup>
Temperature Controllers (RTD, Thermocouples) <sup>3</sup> – Electrical Simulation			EURAMET/cg-11
RTD	(-200 to 850) °C	0.15 °C	Simulator Druck

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments <sup>7</sup>
Temperature Controllers (RTD, Thermocouples) <sup>3</sup> – Electrical Simulation			EURAMET/cg-11
Thermocouples			Simulator Druck
Type K	(-270 to 1370) °C	0.13 °C	
Type J	(-210 to 1200) °C	0.14 °C	
Type R	(-50 to 1760) °C	0.25 °C	
Type S	(-50 to 1740) °C	0.25 °C	
Type T	(-270 to 400) °C	0.18 °C	
Type N	(-270 to 1300) °C	0.24 °C	
pH Meters, Ion Meters and Titrators <sup>3</sup> –			
Electrical Simulation (Voltage)	(-2000 to -700) mV (-700 to 0) mV (0 to 400) mV (400 to 2000) mV	0.21 mV 0.15 mV 0.17 mV 0.21 mV	Simulator Metrohm
Electrical Simulation Temperature – Resistance (PT100, PT1000)	Up to 100 °C	0.06 °C	Resistance decade box 1040, Time Electronics

### III. Fluid Quantities

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments <sup>7</sup>
Volume Measurements Gravimetric Method <sup>3,6</sup> –			Balance
Pipettes	(2 to 10) µl <sup>5</sup>	0.13 µl	ISO 8655-1
Burettes	20 µl	0.13 µl	ISO 8655-2
Dispensers	50 µl	0.15 µl	ISO 8655-3
Dilators	100 µl	0.19 µl	ISO 8655-4
	200 µl	0.21 µl	ISO 8655-5
	500 µl	0.28 µl	ISO 8655-6
	1 ml	0.45 µl	
	2 ml	1.2 µl	
	5 ml	2.3 µl	
	10 ml	2.9 µl	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments <sup>7</sup>
Volume Measurements Gravimetric Method <sup>3,6</sup> – (cont)			Balance
Pipettes	20 ml	5.7 µl	ISO 8655-1
Burettes	50 ml	9.1 µl	ISO 8655-2
Dispensers	100 ml	14 µl	ISO 8655-3
Dilators			ISO 8655-4
			ISO 8655-5
			ISO 8655-6

#### IV. Mechanical

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments <sup>7</sup>
Scales & Balances <sup>3</sup> –	Up to 5 kg R = 0.000001 g R ≤ 0.001 g R > 0.001 g	4R 2R 0.8R	Weights class E2 (OIML R111) OIML R76 USP 41
Titration <sup>3</sup> – Spindle Travel	Up to 10 mm (10 to 50) mm (50 to 100) mm	0.007 mm 0.010 mm 0.012 mm	Dial indicator

#### V. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2,8</sup> (±)	Comments <sup>7</sup>
Temperature Measuring Instruments & Devices <sup>3</sup>			ISO 1770 ISO 1771 ASTM E1 ASTM E644 ASTM E1137 ASTM E220
Thermocouples	(-50 to -35) °C (-35 to 0) °C (0 to 100) °C (100 to 300) °C	0.25 °C 0.16 °C 0.19 °C 0.23 °C	
RTD (PRT) Probes	(-50 to 0) °C (0 to 20) °C (20 to 100) °C (100 to 300) °C	0.08 °C 0.05 °C 0.07 °C 0.10 °C	SPRT, RTD probes:

Parameter/Equipment	Range	CMC <sup>2, 8</sup> (±)	Comments <sup>7</sup>
Temperature Measuring Instruments & Devices <sup>3</sup> (cont)			ISO 1770 ISO 1771 ASTM E1 ASTM E644 ASTM E1137 ASTM E220
Liquid-In-Glass Thermometers	(-35 to 0) °C (0 to 20) °C (20 to 100) °C	0.09 °C 0.085 °C 0.095 °C	
Mechanical & Electrical Indicators	Up to 100) °C (100 to 300) °C	0.19 °C 0.23 °C	SPRT, RTD probes:
Temperature – Measure <sup>3</sup>			
Liquid Baths, Heaters	Up to 20 °C (20 to 200) °C (200 to 300) °C	0.21 °C 0.24 °C 0.29 °C	SPRT, RTD probes Thermocouples
Climatic Chambers <sup>3</sup> (Furnaces, Freezers, Incubators, Refrigerators, Vehicles for Food Transportation <sup>10</sup> ) Including Uniformity	(-50 to 0) °C (0 to 20) °C (20 to 200) °C (200 to 300) °C (300 to 600) °C (600 to 800) °C	0.18 °C 0.19 °C 0.26 °C 1.2 °C 2.7 °C 4.4 °C	DKD-R 5-7 Thermocouples, temperature loggers, probe PT100
Humidity – Measure <sup>3</sup>			
Climatic Chambers	(25 to 75) % RH	1.5 % RH	Humidity logger

## VI. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2, 8</sup> (±)	Comments <sup>7</sup>
Rotational Speed, RPM <sup>3</sup>	Up to 300 RPM (300 to 600) RPM (600 to 1200) RPM (1200 to 48 000) RPM	0.20 RPM 0.28 RPM 1.5 RPM 11 RPM	Optical tachometer

<sup>1</sup> This laboratory offers commercial calibration and field calibration services.

- <sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- <sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- <sup>4</sup> In the statement of CMC,  $R$  represents the resolution of the unit under test.
- <sup>5</sup> Measurement for nominal value of more than 10  $\mu\text{l}$  (note a of the table 1 of ISO 8655-6)
- <sup>6</sup> Calibration by means of balances calibrated by Dr. Golik.
- <sup>7</sup> Calibration can be also performed to manufacturer or specific customer requirements.
- <sup>8</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.
- <sup>9</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.
- <sup>10</sup> IS 1291 Israeli standard (Hebrew version) Vehicle for food products transport in a controlled temperature.



## *Accredited Laboratory*

A2LA has accredited

### **DR. GOLIK – SCIENTIFIC SOLUTIONS**

*Tel-Aviv, ISRAEL*

for technical competence in the field of

### **Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system *(refer to joint ISO-ILAC-IAF Communiqué dated April 2017)*.



Presented this 7<sup>th</sup> day of October 2020.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 5874.01  
Valid to November 30, 2022  
Revised October 8, 2020

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*